

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051920\
 Data File : VV016114.D
 Acq On : 19 May 2020 11:17
 Operator : SY/MD
 Sample : VV0519WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK34

Quant Time: May 20 01:40:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 20 01:20:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	152246	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	148465	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	69479	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	39263	4.75	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.00%
7) Chloroethane-d5	1.58	69	40414	5.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.20%
11) 1,1-Dichloroethene-d2	2.12	63	59713	3.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.20%
20) 2-Butanone-d5	3.96	46	127362	54.06	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.12%
24) Chloroform-d	4.37	84	93581	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.60%
26) 1,2-Dichloroethane-d4	5.06	65	54721	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.80%
32) Benzene-d6	5.07	84	178690	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
36) 1,2-Dichloropropane-d6	6.09	67	57545	5.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.80%
41) Toluene-d8	7.34	98	154919	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
45) trans-1,3-Dichloropropene-	7.64	79	20114	4.75	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.00%
48) 2-Hexanone-d5	8.12	63	98773	53.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.26%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	42092	5.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	59164	5.25	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV051920\
Data File : VV016114.D
Acq On : 19 May 2020 11:17
Operator : SY/MD
Sample : VV0519WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK34

Quant Time: May 20 01:40:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 20 01:20:41 2020
Response via : Initial Calibration

