

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051920\
 Data File : VV016120.D
 Acq On : 19 May 2020 14:19
 Operator : SY/MD
 Sample : L2661-10
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: May 20 02:25:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 20 01:20:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	142756	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	144315	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	67120	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	35379	4.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.40%
7) Chloroethane-d5	1.58	69	38424	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.60%
11) 1,1-Dichloroethene-d2	2.12	63	56453	3.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.80%
20) 2-Butanone-d5	3.97	46	131613	59.58	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.16%
24) Chloroform-d	4.38	84	86994	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	5.06	65	52029	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.20%
32) Benzene-d6	5.08	84	164629	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
36) 1,2-Dichloropropane-d6	6.10	67	53999	5.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.00%
41) Toluene-d8	7.34	98	143401	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
45) trans-1,3-Dichloropropene-	7.65	79	18678	4.54	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.80%
48) 2-Hexanone-d5	8.12	63	98744	54.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.28%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	39498	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	56492	5.19	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.80%


Target Compounds					Ovalue
25) Chloroform	4.40	83	5796	0.315 ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051920\
Data File : VV016120.D
Acq On : 19 May 2020 14:19
Operator : SY/MD
Sample : L2661-10
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: May 20 02:25:30 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 20 01:20:41 2020
Response via : Initial Calibration





#25
 Chloroform
 Concen: 0.315 ug/L
 RT: 4.40 min Scan# 1030
 Delta R.T. -0.00 min
 Lab File: VV016120.D
 Acq: 19 May 2020 14:19

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Tot Ion: 83 Resp: 5796
 Ion Ratio Lower Upper
 83 100
 85 68.8 45.7 84.9

