

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051923\
 Data File : VV031145.D
 Acq On : 20 May 2023 07:37
 Operator : SY/MD
 Sample : 02865-19
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBX74

Quant Time: May 22 00:15:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat May 20 01:55:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	246153	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	235811	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	106620	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	87036	41.373	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.740%
7) Chloroethane-d5	1.532	69	71940	44.391	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.780%
11) 1,1-Dichloroethene-d2	2.063	65	41175	41.739	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.480%
21) 2-Butanone-d5	3.793	46	106812	95.288	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.290%
24) Chloroform-d	4.256	84	171824	46.170	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.340%
26) 1,2-Dichloroethane-d4	4.947	65	122376	48.662	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.320%
32) Benzene-d6	4.966	84	307799	48.149	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.300%
36) 1,2-Dichloropropane-d6	5.995	67	94745	49.534	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.060%
41) Toluene-d8	7.249	98	266189	44.615	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.220%
43) trans-1,3-Dichloroprop...	7.558	79	42759	44.557	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.120%
47) 2-Hexanone-d5	8.030	63	66243	88.825	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.830%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	130920	45.635	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.280%
66) 1,2-Dichlorobenzene-d4	11.567	152	104410	54.495	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.980%
Target Compounds						
					Qvalue	
3) Chloromethane	1.217	50	10147	5.832	ug/L	97
6) Bromomethane	1.487	94	2212	2.540	ug/L	99
13) Acetone	2.117	43	19015	13.048	ug/L	98
14) Carbon disulfide	2.246	76	11653	4.677	ug/L #	95
34) Trichloroethene	5.847	95	1380	0.912	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051923\
Data File : VV031145.D
Acq On : 20 May 2023 07:37
Operator : SY/MD
Sample : 02865-19
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBX74

Quant Time: May 22 00:15:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 20 01:55:03 2023
Response via : Initial Calibration

