

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052020\
 Data File : VV016134.D
 Acq On : 20 May 2020 16:15
 Operator : SY/MD
 Sample : VIBLK41
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK41

Quant Time: May 21 03:43:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 21 01:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	140914	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	142431	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	65075	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	31842	4.17	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.40%
7) Chloroethane-d5	1.58	69	36359	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.13	63	57452	3.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.80%
20) 2-Butanone-d5	3.93	46	121537	55.73	ug/L	-0.06
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.46%
24) Chloroform-d	4.38	84	81755	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	5.06	65	48298	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
32) Benzene-d6	5.08	84	163643	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	6.10	67	52825	5.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.20%
41) Toluene-d8	7.34	98	141106	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
45) trans-1,3-Dichloropropene-	7.65	79	17513	4.31	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.20%
48) 2-Hexanone-d5	8.12	63	91788	51.46	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.92%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	39920	5.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.80%
65) 1,2-Dichlorobenzene-d4	11.65	152	55388	5.24	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.80%

Target Compounds					Ovalue
25) Chloroform	4.41	83	8637	0.475 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052020\
Data File : VV016134.D
Acq On : 20 May 2020 16:15
Operator : SY/MD
Sample : VIBLK41
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK41

Quant Time: May 21 03:43:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 21 01:36:12 2020
Response via : Initial Calibration

