

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
 Data File : VV021339.D
 Acq On : 20 May 2021 10:02
 Operator : SY/MD
 Sample : VSTD0.0561
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.05161

Quant Time: May 21 05:00:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 21 05:00:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	5.638	114	29651	0.50 ug/L	0.00
5) Chlorobenzene-d5	8.863	117	27239	0.50 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.243	152	14518	0.50 ug/L	-0.02
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.302	65	1266	0.05 ug/L	0.00
4) 1,2-Dichloroethane-d4	5.066	65	789	0.04 ug/L	0.02
7) 1,2-Dichloropropane-d6	6.068	67	1062	0.04 ug/L	0.00
8) Toluene-d8	7.332	98	3188	0.06 ug/L	0.01
10) 1,1,2,2-Tetrachloroeth...	10.215	84	617	0.04 ug/L	0.00
Target Compounds					
					Qvalue
3) Vinyl chloride	1.302	62	1228	0.03 ug/L	84
6) Trichloroethene	5.925	95	1054	0.04 ug/L	81
9) 1,2-Dibromoethane	8.368	107	625	0.04 ug/L #	93
12) 1,2,3-Trichloropropane	10.267	75	511	0.03 ug/L	94
13) 1,2-Dibromo-3-chloropr...	12.426	75	105	0.05 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
Data File : VV021339.D
Acq On : 20 May 2021 10:02
Operator : SY/MD
Sample : VSTD0.0561
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.05161

Quant Time: May 21 05:00:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 21 05:00:22 2021
Response via : Initial Calibration

