

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
 Data File : VV021343.D
 Acq On : 20 May 2021 12:04
 Operator : SY/MD
 Sample : VSTD00265
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002165

Quant Time: May 21 05:01:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 21 05:00:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	28785	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	27534	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	14546	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	27720	1.12	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	28857	1.56	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	40265	1.66	ug/L	0.00
8) Toluene-d8	7.317	98	111946	2.13	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	22216	1.46	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	46131	1.10	ug/L	96
6) Trichloroethene	5.925	95	40913	1.61	ug/L	97
9) 1,2-Dibromoethane	8.353	107	27013	1.82	ug/L #	97
12) 1,2,3-Trichloropropane	10.267	75	22112	1.49	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.409	75	3520	1.61	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
Data File : VV021343.D
Acq On : 20 May 2021 12:04
Operator : SY/MD
Sample : VSTD00265
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002165

Quant Time: May 21 05:01:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 21 05:00:22 2021
Response via : Initial Calibration

