

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
 Data File : VV021345.D
 Acq On : 20 May 2021 13:20
 Operator : SY/MD
 Sample : VV0520WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK1962

Quant Time: May 21 05:08:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 21 05:04:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	29465	0.50	ug/L	-0.02
5) Chlorobenzene-d5	8.863	117	26773	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	14316	0.50	ug/L	-0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	7813	0.43	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	86.00%
4) 1,2-Dichloroethane-d4	5.042	65	7668	0.50	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.00%
7) 1,2-Dichloropropane-d6	6.068	67	10674	0.52	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	104.00%
8) Toluene-d8	7.317	98	29851	0.51	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.00%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	5594	0.49	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	98.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
Data File : VV021345.D
Acq On : 20 May 2021 13:20
Operator : SY/MD
Sample : VV0520WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK1962

Quant Time: May 21 05:08:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 21 05:04:41 2021
Response via : Initial Calibration

