

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
 Data File : VV021347.D
 Acq On : 20 May 2021 14:34
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5167

Quant Time: May 21 05:08:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 21 05:04:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	30207	0.50	ug/L	-0.02
5) Chlorobenzene-d5	8.863	117	27432	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	14853	0.50	ug/L	-0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.302	65	8200	0.45	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	90.00%		
4) 1,2-Dichloroethane-d4	5.042	65	7926	0.51	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	102.00%		
7) 1,2-Dichloropropane-d6	6.068	67	10851	0.52	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	104.00%		
8) Toluene-d8	7.316	98	31043	0.52	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	104.00%		
10) 1,1,2,2-Tetrachloroeth...	10.215	84	5668	0.49	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	98.00%		
Target Compounds					Qvalue	
3) Vinyl chloride	1.302	62	12363	0.49	ug/L	99
6) Trichloroethene	5.925	95	10822	0.50	ug/L	99
9) 1,2-Dibromoethane	8.353	107	6937	0.50	ug/L	97
12) 1,2,3-Trichloropropane	10.267	75	5641	0.49	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.409	75	874	0.46	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
Data File : VV021347.D
Acq On : 20 May 2021 14:34
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5167

Quant Time: May 21 05:08:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 21 05:04:41 2021
Response via : Initial Calibration

