

Method Path : Z:\voasrv\HPCHEM1\MSVOA_V\Method\
 Method File : SFAMVSIM052021.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri May 21 05:04:41 2021
 Response Via : Initial Calibration

Calibration Files

0.5 =VV021341.D 0.05=VV021339.D 0.1 =VV021340.D 1.0 =VV021342.D 2 =VV021343.D

	Compound	0.5	0.05	0.1	1.0	2	Avg	%RSD

1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.281	0.427	0.325	0.251	0.241	0.305	24.80
3) T	Vinyl chloride	0.442	0.414	0.425	0.408	0.401	0.418	3.86
4) S	1,2-Dichloroetha...	0.276	0.266	0.235	0.263	0.251	0.258	6.12
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.406	0.387	0.387	0.416	0.371	0.393	4.46
7) S	1,2-Dichloroprop...	0.409	0.390	0.367	0.386	0.366	0.383	4.66
8) S	Toluene-d8	1.123	1.170	1.019	1.107	1.016	1.087	6.23
9) T	1,2-Dibromoethane	0.273	0.229	0.256	0.265	0.245	0.254	6.71
10) S	1,1,2,2-Tetrachl...	0.222	0.227	0.198	0.208	0.202	0.211	5.86
11) I	1,4-Dichlorobenzen...	-----ISTD-----						
12)	1,2,3-Trichlorop...	0.419	0.352	0.393	0.405	0.380	0.390	6.56
13) T	1,2-Dibromo-3-ch...	0.065	0.072	0.062	0.063	0.060	0.064	7.33

(#) = Out of Range