

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
 Data File : VV021340.D
 Acq On : 20 May 2021 10:39
 Operator : SY/MD
 Sample : VSTD0.162
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.1162

Quant Time: May 21 05:00:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 21 05:00:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	5.615	114	33564	0.50 ug/L	-0.02
5) Chlorobenzene-d5	8.864	117	30515	0.50 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	16305	0.50 ug/L	-0.02
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.303	65	2180	0.08 ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	1575	0.07 ug/L	0.00
7) 1,2-Dichloropropane-d6	6.069	67	2238	0.08 ug/L	0.00
8) Toluene-d8	7.317	98	6216	0.11 ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	1210	0.07 ug/L	0.00
Target Compounds					
					Qvalue
3) Vinyl chloride	1.303	62	2856	0.06 ug/L	94
6) Trichloroethene	5.925	95	2361	0.08 ug/L	92
9) 1,2-Dibromoethane	8.369	107	1562	0.09 ug/L #	98
12) 1,2,3-Trichloropropane	10.267	75	1283	0.08 ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.427	75	201	0.08 ug/L #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
Data File : VV021340.D
Acq On : 20 May 2021 10:39
Operator : SY/MD
Sample : VSTD0.162
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.1162

Quant Time: May 21 05:00:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 21 05:00:22 2021
Response via : Initial Calibration

