

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
 Data File : VV021344.D
 Acq On : 20 May 2021 12:35
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV166

Quant Time: May 21 05:08:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 21 05:04:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	30323	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	27417	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	14878	0.50	ug/L	-0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	7932	0.43	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 86.00%			
4) 1,2-Dichloroethane-d4	5.042	65	7778	0.50	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 100.00%			
7) 1,2-Dichloropropane-d6	6.068	67	10758	0.51	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 102.00%			
8) Toluene-d8	7.317	98	30548	0.51	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 102.00%			
10) 1,1,2,2-Tetrachloroeth...	10.216	84	5787	0.50	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 100.00%			
Target Compounds						
						Qvalue
3) Vinyl chloride	1.303	62	12611	0.50	ug/L	100
6) Trichloroethene	5.925	95	11107	0.51	ug/L	99
9) 1,2-Dibromoethane	8.353	107	7239	0.52	ug/L	96
12) 1,2,3-Trichloropropane	10.267	75	5970	0.51	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.409	75	917	0.48	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052021\
Data File : VV021344.D
Acq On : 20 May 2021 12:35
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV166

Quant Time: May 21 05:08:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM052021.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 21 05:04:41 2021
Response via : Initial Calibration

