

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052120\
 Data File : VV016147.D
 Acq On : 21 May 2020 12:47
 Operator : SY/MD
 Sample : L2725-15
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: May 22 02:51:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 22 01:50:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	147235	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	142573	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	66821	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	31231	3.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.20%
7) Chloroethane-d5	1.58	69	35292	4.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.20%
11) 1,1-Dichloroethene-d2	2.12	63	56855	3.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.20%
20) 2-Butanone-d5	3.95	46	89461	39.26	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.52%
24) Chloroform-d	4.38	84	85289	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
26) 1,2-Dichloroethane-d4	5.06	65	44304	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.60%
32) Benzene-d6	5.08	84	160807	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
36) 1,2-Dichloropropane-d6	6.09	67	50402	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.34	98	139708	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.20%
45) trans-1,3-Dichloropropene-	7.65	79	14820	3.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.00%
48) 2-Hexanone-d5	8.12	63	68129	38.16	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.32%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	31998	4.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.80%
65) 1,2-Dichlorobenzene-d4	11.65	152	51899	4.79	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.80%

Target Compounds

					Ovalue
13) Acetone	2.24	43	9938m	6.612	ug/L
16) Methylene chloride	2.53	84	1579	0.163	ug/L
42) Toluene	7.41	91	11706	0.308	ug/L
55) m,p-xylene	9.16	106	2262	0.141	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052120\
Data File : VV016147.D
Acq On : 21 May 2020 12:47
Operator : SY/MD
Sample : L2725-15
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: May 22 02:51:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 22 01:50:48 2020
Response via : Initial Calibration

