

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052120\
 Data File : VV016148.D
 Acq On : 21 May 2020 13:10
 Operator : SY/MD
 Sample : L2725-18
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9B35

Quant Time: May 22 02:58:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 22 01:50:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	145026	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	140126	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	64307	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	31347	3.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.60%
7) Chloroethane-d5	1.58	69	35473	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.20%
11) 1,1-Dichloroethene-d2	2.13	63	55746	3.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.80%
20) 2-Butanone-d5	3.94	46	92781	41.34	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.68%
24) Chloroform-d	4.38	84	86480	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
26) 1,2-Dichloroethane-d4	5.06	65	43896	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
32) Benzene-d6	5.08	84	161093	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	6.10	67	50929	5.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.00%
41) Toluene-d8	7.34	98	141337	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
45) trans-1,3-Dichloropropene-	7.65	79	14638	3.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.40%
48) 2-Hexanone-d5	8.12	63	70501	40.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.36%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	31917	4.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	52224	5.00	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.00%

Target Compounds					Ovalue
13) Acetone	2.23	43	11911	8.045 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052120\
Data File : VV016148.D
Acq On : 21 May 2020 13:10
Operator : SY/MD
Sample : L2725-18
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9B35

Quant Time: May 22 02:58:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 22 01:50:48 2020
Response via : Initial Calibration

