

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052120\
 Data File : VV016160.D
 Acq On : 21 May 2020 17:46
 Operator : SY/MD
 Sample : L2731-04
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AW9

Quant Time: May 22 05:11:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 22 01:50:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.65	114	137327	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	131794	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	60902	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	28714	3.85	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.00%
7) Chloroethane-d5	1.58	69	32996	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.40%
11) 1,1-Dichloroethene-d2	2.13	63	53979	3.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.40%
20) 2-Butanone-d5	3.94	46	89654	42.19	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.38%
24) Chloroform-d	4.38	84	80369	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.06	65	42383	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
32) Benzene-d6	5.08	84	154269	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
36) 1,2-Dichloropropane-d6	6.10	67	48627	5.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.60%
41) Toluene-d8	7.34	98	134065	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
45) trans-1,3-Dichloropropene-	7.65	79	13793	3.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.40%
48) 2-Hexanone-d5	8.12	63	67350	40.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.62%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	30794	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.40%
65) 1,2-Dichlorobenzene-d4	11.65	152	49457	5.00	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052120\
Data File : VV016160.D
Acq On : 21 May 2020 17:46
Operator : SY/MD
Sample : L2731-04
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AW9

Quant Time: May 22 05:11:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 22 01:50:48 2020
Response via : Initial Calibration

