

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052120\
 Data File : VV016164.D
 Acq On : 21 May 2020 19:18
 Operator : SY/MD
 Sample : L2731-08
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AX3

Quant Time: May 22 05:26:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 22 01:50:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	116552	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	114147	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	49173	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	28021	4.43	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.60%
7) Chloroethane-d5	1.58	69	32767	5.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.40%
11) 1,1-Dichloroethene-d2	2.13	63	52709	3.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.60%
20) 2-Butanone-d5	3.94	46	102447	56.80	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.60%
24) Chloroform-d	4.38	84	81334	5.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	117.60%
26) 1,2-Dichloroethane-d4	5.06	65	44881	5.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.20%
32) Benzene-d6	5.08	84	152477	5.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.20%
36) 1,2-Dichloropropane-d6	6.10	67	50015	6.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	120.60%
41) Toluene-d8	7.34	98	132362	5.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.00%
45) trans-1,3-Dichloropropene-	7.65	79	14041	4.32	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.40%
48) 2-Hexanone-d5	8.12	63	79024	55.28	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.56%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	33318	5.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	114.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	49968	6.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	125.20%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052120\
Data File : VV016164.D
Acq On : 21 May 2020 19:18
Operator : SY/MD
Sample : L2731-08
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AX3

Quant Time: May 22 05:26:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 22 01:50:48 2020
Response via : Initial Calibration

