

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052121\
 Data File : VV021355.D
 Acq On : 21 May 2021 13:36
 Operator : SY/MD
 Sample : VV0521WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK170

Quant Time: May 22 01:39:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 21 12:12:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	294676	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	277009	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	153328	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	87995	4.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.80%
7) Chloroethane-d5	1.565	69	72592	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.40%
11) 1,1-Dichloroethene-d2	2.101	63	112342	3.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.80%
20) 2-Butanone-d5	3.937	46	251935	49.05	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.10%
24) Chloroform-d	4.352	84	185167	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	5.037	65	91993	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
32) Benzene-d6	5.053	84	390304	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
36) 1,2-Dichloropropane-d6	6.072	67	121957	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	7.320	98	364558	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
43) trans-1,3-Dichloroprop...	7.625	79	52536	4.91	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.20%
46) 2-Hexanone-d5	8.095	63	171318	44.88	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.76%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	67490	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.60%
66) 1,2-Dichlorobenzene-d4	11.622	152	149782	4.81	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.20%
Target Compounds						Qvalue
16) Methylene chloride	2.507	84	12403	0.54	ug/L	97
25) Chloroform	4.384	83	6230	0.17	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052121\
Data File : VV021355.D
Acq On : 21 May 2021 13:36
Operator : SY/MD
Sample : VV0521WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK170

Quant Time: May 22 01:39:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 21 12:12:53 2021
Response via : Initial Calibration

