

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052121\
 Data File : VV021364.D
 Acq On : 21 May 2021 17:22
 Operator : SY/MD
 Sample : M2494-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB421

Quant Time: May 22 01:40:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 21 12:12:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	281430	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	262921	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	142621	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	80125	4.37	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.40%
7) Chloroethane-d5	1.568	69	66944	4.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.20%
11) 1,1-Dichloroethene-d2	2.108	63	124642	4.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.20%
20) 2-Butanone-d5	3.896	46	227435	46.36	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.72%
24) Chloroform-d	4.352	84	165180	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.80%
26) 1,2-Dichloroethane-d4	5.037	65	81919	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
32) Benzene-d6	5.053	84	345242	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
36) 1,2-Dichloropropane-d6	6.072	67	109323	4.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.00%
41) Toluene-d8	7.320	98	317765	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.80%
43) trans-1,3-Dichloroprop...	7.628	79	44787	4.41	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.20%
46) 2-Hexanone-d5	8.088	63	153744	42.43	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.86%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	61370	4.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.00%
66) 1,2-Dichlorobenzene-d4	11.625	152	131246	4.53	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.60%
Target Compounds						
12) 1,1-Dichloroethene	2.117	96	43485	2.80	ug/L #	67
19) 1,1-Dichloroethane	3.195	63	14653	0.44	ug/L	97
25) Chloroform	4.381	83	4745	0.13	ug/L	85
29) 1,1,1-Trichloroethane	4.613	97	83612	2.59	ug/L	98
34) Trichloroethene	5.921	95	1975	0.09	ug/L	85
42) Toluene	7.403	91	6252	0.08	ug/L	93
47) Tetrachloroethene	7.979	164	87563	4.78	ug/L	97
48) 2-Hexanone	8.146	43	40379	4.84	ug/L #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052121\
Data File : VV021364.D
Acq On : 21 May 2021 17:22
Operator : SY/MD
Sample : M2494-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB421

Quant Time: May 22 01:40:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 21 12:12:53 2021
Response via : Initial Calibration

