

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052121\
 Data File : VW021367.D
 Acq On : 21 May 2021 18:36
 Operator : SY/MD
 Sample : M2494-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB424

Quant Time: May 22 01:40:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 21 12:12:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	277191	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	253851	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	138366	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	78497	4.35	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.00%
7) Chloroethane-d5	1.568	69	65991	4.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.20%
11) 1,1-Dichloroethene-d2	2.111	63	215595	7.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	142.40%#
20) 2-Butanone-d5	3.895	46	223078	46.17	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.34%
24) Chloroform-d	4.352	84	167030	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.00%
26) 1,2-Dichloroethane-d4	5.037	65	80739	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
32) Benzene-d6	5.053	84	344587	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
36) 1,2-Dichloropropane-d6	6.072	67	108133	4.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.40%
41) Toluene-d8	7.320	98	319901	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
43) trans-1,3-Dichloroprop...	7.628	79	43272	4.41	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.20%
46) 2-Hexanone-d5	8.088	63	148617	42.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.96%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	60891	4.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.20%
66) 1,2-Dichlorobenzene-d4	11.625	152	130849	4.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.00%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.117	96	236651	15.47	ug/L #	84
19) 1,1-Dichloroethane	3.191	63	127469	3.87	ug/L	99
22) cis-1,2-Dichloroethene	3.921	96	4771	0.23	ug/L	89
25) Chloroform	4.381	83	5883	0.17	ug/L	89
27) 1,2-Dichloroethane	5.149	62	6008	0.32	ug/L #	85
29) 1,1,1-Trichloroethane	4.609	97	111940	3.59	ug/L	97
34) Trichloroethene	5.921	95	6171	0.28	ug/L	95
42) Toluene	7.400	91	6367	0.08	ug/L	87
47) Tetrachloroethene	7.976	164	229238	12.96	ug/L	97
48) 2-Hexanone	8.146	43	41662	5.17	ug/L #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052121\
Data File : VV021367.D
Acq On : 21 May 2021 18:36
Operator : SY/MD
Sample : M2494-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB424

Quant Time: May 22 01:40:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 21 12:12:53 2021
Response via : Initial Calibration

