

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052121\
 Data File : VV021369.D
 Acq On : 21 May 2021 19:25
 Operator : SY/MD
 Sample : M2494-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB426

Quant Time: May 22 01:41:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 21 12:12:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	277506	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	255978	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	138538	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	76007	4.21	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.20%
7) Chloroethane-d5	1.568	69	64679	4.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.40%
11) 1,1-Dichloroethene-d2	2.105	63	107974	3.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.20%
20) 2-Butanone-d5	3.905	46	218509	45.17	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.34%
24) Chloroform-d	4.352	84	159537	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.00%
26) 1,2-Dichloroethane-d4	5.037	65	78070	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
32) Benzene-d6	5.053	84	332799	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
36) 1,2-Dichloropropane-d6	6.072	67	105944	4.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.60%
41) Toluene-d8	7.320	98	306720	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
43) trans-1,3-Dichloroprop...	7.628	79	41519	4.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.00%
46) 2-Hexanone-d5	8.091	63	147108	41.70	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.40%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	59173	4.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.20%
66) 1,2-Dichlorobenzene-d4	11.625	152	125607	4.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.20%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.117	96	23057	1.51	ug/L #	36
19) 1,1-Dichloroethane	3.195	63	12124	0.37	ug/L	95
25) Chloroform	4.378	83	4295	0.12	ug/L	90
29) 1,1,1-Trichloroethane	4.613	97	6542	0.21	ug/L	94
47) Tetrachloroethene	7.979	164	21647	1.21	ug/L	96
48) 2-Hexanone	8.143	43	41054	5.05	ug/L #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052121\
Data File : VV021369.D
Acq On : 21 May 2021 19:25
Operator : SY/MD
Sample : M2494-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB426

Quant Time: May 22 01:41:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 21 12:12:53 2021
Response via : Initial Calibration

