

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052121\
 Data File : VV021370.D
 Acq On : 21 May 2021 19:49
 Operator : SY/MD
 Sample : M2494-14
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB429

Quant Time: May 22 01:41:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 21 12:12:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	278550	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	259657	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	138258	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	81430	4.49	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	89.80%	
7) Chloroethane-d5	1.568	69	69478	4.73	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.60%	
11) 1,1-Dichloroethene-d2	2.108	63	104672	3.44	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	68.80%	
20) 2-Butanone-d5	3.899	46	235321	48.46	ug/L	-0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	=	96.92%	
24) Chloroform-d	4.352	84	172500	4.57	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.40%	
26) 1,2-Dichloroethane-d4	5.037	65	83625	4.72	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.40%	
32) Benzene-d6	5.053	84	359906	4.66	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.20%	
36) 1,2-Dichloropropane-d6	6.072	67	112696	4.75	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.00%	
41) Toluene-d8	7.320	98	328724	4.54	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.80%	
43) trans-1,3-Dichloroprop...	7.628	79	43743	4.36	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.20%	
46) 2-Hexanone-d5	8.091	63	156681	43.79	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	87.58%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	64067	4.44	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	88.80%	
66) 1,2-Dichlorobenzene-d4	11.625	152	135132	4.81	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	96.20%	
Target Compounds						
25) Chloroform	4.375	83	3657	0.11	ug/L #	72
42) Toluene	7.397	91	6830	0.08	ug/L	97
48) 2-Hexanone	8.143	43	43377	5.26	ug/L #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052121\
Data File : VV021370.D
Acq On : 21 May 2021 19:49
Operator : SY/MD
Sample : M2494-14
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB429

Quant Time: May 22 01:41:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 21 12:12:53 2021
Response via : Initial Calibration

