

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052223\
 Data File : VW031168.D
 Acq On : 22 May 2023 16:07
 Operator : SY/MD
 Sample : 02861-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0FL8

Quant Time: May 23 04:52:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 23 04:42:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	231576	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	225800	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	96586	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	67219	33.965	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	67.920%		
7) Chloroethane-d5	1.536	69	58137	38.132	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	76.260%		
11) 1,1-Dichloroethene-d2	2.063	65	34872	37.575	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	75.140%		
21) 2-Butanone-d5	3.799	46	100187	95.004	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	95.000%		
24) Chloroform-d	4.259	84	151858	43.373	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.740%		
26) 1,2-Dichloroethane-d4	4.950	65	109854	46.433	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.860%		
32) Benzene-d6	4.970	84	258350	42.205	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	84.420%		
36) 1,2-Dichloropropane-d6	5.998	67	80945	44.195	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	88.400%		
41) Toluene-d8	7.249	98	225101	39.401	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	78.800%#		
43) trans-1,3-Dichloroprop...	7.558	79	38858	42.288	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	84.580%		
47) 2-Hexanone-d5	8.034	63	63243	88.562	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	88.560%		
56) 1,1,2,2-Tetrachloroeth...	10.159	84	97882	35.632	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	71.260%		
66) 1,2-Dichlorobenzene-d4	11.567	152	83944	48.365	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	96.720%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052223\
Data File : VV031168.D
Acq On : 22 May 2023 16:07
Operator : SY/MD
Sample : 02861-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0FL8

Quant Time: May 23 04:52:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
Quant Title : VOC Analysis
QLast Update : Tue May 23 04:42:12 2023
Response via : Initial Calibration

