

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052223\
 Data File : VW031153.D
 Acq On : 22 May 2023 10:13
 Operator : SY/MD
 Sample : 02866-05DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBX80DL

Quant Time: May 23 04:49:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 23 04:42:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	259797	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	253526	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	121465	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	83239	37.490	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.980%
7) Chloroethane-d5	1.535	69	73262	42.832	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.660%
11) 1,1-Dichloroethene-d2	2.066	65	42394	40.718	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.440%
21) 2-Butanone-d5	3.802	46	99968	84.499	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.500%
24) Chloroform-d	4.255	84	171244	43.597	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.200%
26) 1,2-Dichloroethane-d4	4.950	65	121175	45.654	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.300%
32) Benzene-d6	4.969	84	303271	44.126	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.260%
36) 1,2-Dichloropropane-d6	5.995	67	93301	45.370	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.740%
41) Toluene-d8	7.249	98	266289	41.513	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.020%
43) trans-1,3-Dichloroprop...	7.561	79	44887	43.506	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.020%
47) 2-Hexanone-d5	8.033	63	63304	78.953	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.950%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	129969	42.138	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.280%
66) 1,2-Dichlorobenzene-d4	11.567	152	108838	49.863	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.720%
Target Compounds						
					Qvalue	
34) Trichloroethene	5.837	95	98420	60.475	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052223\
Data File : VV031153.D
Acq On : 22 May 2023 10:13
Operator : SY/MD
Sample : 02866-05DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBX80DL

Quant Time: May 23 04:49:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
Quant Title : VOC Analysis
QLast Update : Tue May 23 04:42:12 2023
Response via : Initial Calibration

