

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052223\
 Data File : VW031156.D
 Acq On : 22 May 2023 11:25
 Operator : SY/MD
 Sample : 02866-08DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBX83DL

Quant Time: May 23 04:49:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 23 04:42:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	259716	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	251154	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	117959	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	79701	35.908	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.820%
7) Chloroethane-d5	1.536	69	69771	40.804	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.600%
11) 1,1-Dichloroethene-d2	2.066	65	40275	38.695	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.380%
21) 2-Butanone-d5	3.799	46	100502	84.977	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.980%
24) Chloroform-d	4.259	84	165947	42.262	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.520%
26) 1,2-Dichloroethane-d4	4.950	65	117236	44.184	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.360%
32) Benzene-d6	4.969	84	294195	43.209	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.420%
36) 1,2-Dichloropropane-d6	5.998	67	90945	44.642	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.280%
41) Toluene-d8	7.249	98	263437	41.456	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.920%
43) trans-1,3-Dichloroprop...	7.558	79	43843	42.896	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.800%
47) 2-Hexanone-d5	8.034	63	59502	74.912	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.910%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	125452	41.058	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.120%
66) 1,2-Dichlorobenzene-d4	11.567	152	106081	50.045	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.080%
Target Compounds						
					Qvalue	
20) cis-1,2-Dichloroethene	3.831	96	5781	3.516	ug/L	94
34) Trichloroethene	5.841	95	74449	46.178	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052223\
Data File : VV031156.D
Acq On : 22 May 2023 11:25
Operator : SY/MD
Sample : 02866-08DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBX83DL

Quant Time: May 23 04:49:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
Quant Title : VOC Analysis
QLast Update : Tue May 23 04:42:12 2023
Response via : Initial Calibration









#20
cis-1,2-Dichloroethene
Concen: 3.516 ug/L
RT: 3.831 min Scan# 865
Delta R.T. 0.010 min
Lab File: VV031156.D
Acq: 22 May 2023 11:25

Instrument :
MSVOA_V
ClientSampleId :
GBX83DL

Tgt Ion: 96 Resp: 5781
Ion Ratio Lower Upper
96 100
61 122.8 89.2 165.6
98 70.0 43.0 80.0



#34
Trichloroethene
Concen: 46.178 ug/L
RT: 5.841 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VV031156.D
Acq: 22 May 2023 11:25

Tgt Ion: 95 Resp: 74449
Ion Ratio Lower Upper
95 100
97 62.6 44.4 82.5
132 102.2 70.2 130.4
130 101.9 71.5 132.7

