

Data File : VV011048.D
Acq On : 24 May 2019 03:11
Operator : SY/MD
Sample : K3016-16
Misc : 4.84G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A51D3

Quant Time: May 24 09:32:33 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM052019S.M
Quant Title : VOC Analysis
QLast Update : Fri May 24 09:03:49 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	92458	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	90749	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	28542	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	18512	17.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.24%
7) Chloroethane-d5	1.56	69	15384	21.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.24%
10) 1,1-Dichloroethene-d2	2.12	63	28012	13.82	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.28%
20) 2-Butanone-d5	3.93	46	35081	96.45	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	192.90%#
24) Chloroform-d	4.40	84	64237	26.82	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	107.28%
26) 1,2-Dichloroethane-d4	5.08	65	41327	29.28	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	117.12%
29) Benzene-d6	5.09	84	123785	24.99	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.96%
33) 1,2-Dichloropropane-d6	6.12	67	45463	28.88	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	115.52%
37) Toluene-d8	7.36	98	94338	20.48	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.92%
38) trans-1,3-Dichloropropene-	7.66	79	15002	23.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.84%
39) 2-Hexanone-d5	8.13	63	28087	95.66	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	191.32%#
48) 1,1,2,2-Tetrachloroethane-	10.26	84	53878	34.28	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	137.12%#
61) 1,2-Dichlorobenzene-d4	11.67	152	29687	25.62	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.48%

Target Compounds					Qvalue
13) Acetone	2.19	43	23168	66.055 ug/L	98
15) Methyl Acetate	2.45	43	3735	5.542 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011048.D
Acq On : 24 May 2019 03:11
Operator : SY/MD
Sample : K3016-16
Misc : 4.84G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A51D3

Quant Time: May 24 09:32:33 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM052019S.M
Quant Title : VOC Analysis
QLast Update : Fri May 24 09:03:49 2019
Response via : Initial Calibration

