

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052323\
 Data File : VW031178.D
 Acq On : 23 May 2023 10:34
 Operator : SY/MD
 Sample : 02861-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0FK6

Quant Time: May 24 01:32:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 24 01:27:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	230580	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	178406	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	112369	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	101327	51.420	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 102.840%	
7) Chloroethane-d5	1.535	69	78427	51.662	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 103.320%	
11) 1,1-Dichloroethene-d2	2.063	65	49836	53.931	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 107.860%	
21) 2-Butanone-d5	3.796	46	100246	95.471	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 95.470%	
24) Chloroform-d	4.259	84	184014	52.785	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 105.560%	
26) 1,2-Dichloroethane-d4	4.950	65	133552	56.693	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 113.380%	
32) Benzene-d6	4.969	84	339297	70.154	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 140.300%#	
36) 1,2-Dichloropropane-d6	5.995	67	101202	69.934	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 139.860%#	
41) Toluene-d8	7.249	98	279992	62.028	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 124.060%#	
43) trans-1,3-Dichloroprop...	7.561	79	45233	62.302	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 124.600%	
47) 2-Hexanone-d5	8.030	63	58247	103.234	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 103.230%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	134034	61.754	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 123.500%#	
66) 1,2-Dichlorobenzene-d4	11.567	152	122120	60.477	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 120.960%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052323\
Data File : VV031178.D
Acq On : 23 May 2023 10:34
Operator : SY/MD
Sample : 02861-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0FK6

Quant Time: May 24 01:32:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 24 01:27:41 2023
Response via : Initial Calibration

