

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052323\
 Data File : VV031179.D
 Acq On : 23 May 2023 10:58
 Operator : SY/MD
 Sample : VV0523WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK343

Quant Time: May 24 01:32:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 24 01:27:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	227214	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	221205	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	108475	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	99895	51.444	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 102.880%	
7) Chloroethane-d5	1.535	69	78406	52.413	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 104.820%	
11) 1,1-Dichloroethene-d2	2.063	65	49154	53.981	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 107.960%	
21) 2-Butanone-d5	3.796	46	101768	98.356	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 98.360%	
24) Chloroform-d	4.259	84	176722	51.444	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 102.880%	
26) 1,2-Dichloroethane-d4	4.950	65	126573	54.526	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 109.060%	
32) Benzene-d6	4.969	84	323801	53.997	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 108.000%	
36) 1,2-Dichloropropane-d6	5.998	67	86047	47.957	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 95.920%	
41) Toluene-d8	7.249	98	255029	45.567	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 91.140%	
43) trans-1,3-Dichloroprop...	7.558	79	45477	50.519	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 101.040%	
47) 2-Hexanone-d5	8.030	63	64005	91.491	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 91.490%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	131158	48.737	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 97.480%	
66) 1,2-Dichlorobenzene-d4	11.567	152	115007	58.999	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 118.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052323\
Data File : VV031179.D
Acq On : 23 May 2023 10:58
Operator : SY/MD
Sample : VV0523WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK343

Quant Time: May 24 01:32:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 24 01:27:41 2023
Response via : Initial Calibration

