

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052323\
 Data File : VW031183.D
 Acq On : 23 May 2023 12:33
 Operator : SY/MD
 Sample : 02861-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0FL1

Quant Time: May 24 01:33:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 24 01:27:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	237233	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	206842	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	114300	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	95162	46.937	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.880%
7) Chloroethane-d5	1.535	69	76499	48.979	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.960%
11) 1,1-Dichloroethene-d2	2.063	65	46382	48.785	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.580%
21) 2-Butanone-d5	3.799	46	94568	87.537	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.540%
24) Chloroform-d	4.256	84	168567	46.998	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.000%
26) 1,2-Dichloroethane-d4	4.950	65	122764	50.652	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.300%
32) Benzene-d6	4.969	84	311477	55.548	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.100%
36) 1,2-Dichloropropane-d6	5.995	67	88990	53.041	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.080%
41) Toluene-d8	7.249	98	261610	49.988	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.980%
43) trans-1,3-Dichloroprop...	7.558	79	40783	48.450	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.900%
47) 2-Hexanone-d5	8.034	63	53358	81.568	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.570%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	123204	48.960	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.920%
66) 1,2-Dichlorobenzene-d4	11.567	152	112267	54.659	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052323\
Data File : VV031183.D
Acq On : 23 May 2023 12:33
Operator : SY/MD
Sample : 02861-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0FL1

Quant Time: May 24 01:33:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 24 01:27:41 2023
Response via : Initial Calibration

