

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052323\
 Data File : VW031180.D
 Acq On : 23 May 2023 11:22
 Operator : SY/MD
 Sample : 02861-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0FK8

Quant Time: May 24 01:33:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 24 01:27:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	212721	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	227509	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	111446	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	101946	56.077	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 112.160%			
7) Chloroethane-d5	1.535	69	76767	54.814	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 109.620%			
11) 1,1-Dichloroethene-d2	2.066	65	47390	55.589	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 111.180%			
21) 2-Butanone-d5	3.796	46	103962	107.322	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 107.320%			
24) Chloroform-d	4.256	84	179048	55.672	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 111.340%			
26) 1,2-Dichloroethane-d4	4.950	65	127220	58.539	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 117.080%			
32) Benzene-d6	4.969	84	327853	53.157	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 106.320%			
36) 1,2-Dichloropropane-d6	5.995	67	85191	46.164	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 92.320%			
41) Toluene-d8	7.249	98	235828	40.968	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 81.940%			
43) trans-1,3-Dichloroprop...	7.558	79	40752	44.016	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 88.040%			
47) 2-Hexanone-d5	8.034	63	63226	87.873	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 87.870%			
56) 1,1,2,2-Tetrachloroeth...	10.159	84	131435	47.486	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 94.980%			
66) 1,2-Dichlorobenzene-d4	11.567	152	116555	58.199	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 116.400%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052323\
Data File : VV031180.D
Acq On : 23 May 2023 11:22
Operator : SY/MD
Sample : 02861-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0FK8

Quant Time: May 24 01:33:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 24 01:27:41 2023
Response via : Initial Calibration

