




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052423\
 Data File : VV031222.D
 Acq On : 24 May 2023 17:57
 Operator : SY/MD
 Sample : 02866-15DL 50X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBX90DL

Manual IntegrationsAPPROVED

Quant Time: May 25 01:16:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 25 01:11:46 2023
 Response via : Initial Calibration

Reviewed By :Krupa Patel 05/25/2023
 Supervised By :Mahesh Dadoda 05/25/2023



TIC: VV031222.D\data.ms

(5) Vinyl chloride (T)

1.092min (-0.196) 0.14 ug/L

response 154

Ion	Exp%	Act%
62.00	100.00	100.00
64.00	33.00	42.80
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052423\
 Data File : VV031222.D
 Acq On : 24 May 2023 17:57
 Operator : SY/MD
 Sample : 02866-15DL 50X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBX90DL

Manual IntegrationsAPPROVED

Quant Time: May 25 01:16:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 25 01:11:46 2023
 Response via : Initial Calibration

Reviewed By :Krupa Patel 05/25/2023
 Supervised By :Mahesh Dadoda 05/25/2023



TIC: VV031222.D\data.ms

(5) Vinyl chloride (T)

1.288min (-0.000) 2.51 ug/L m

response 2859

Ion	Exp%	Act%
62.00	100.00	100.00
64.00	33.00	59.67#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052423\
 Data File : VV031222.D
 Acq On : 24 May 2023 17:57
 Operator : SY/MD
 Sample : 02866-15DL 50X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBX90DL

Manual IntegrationsAPPROVED

Quant Time: May 25 01:16:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 25 01:11:46 2023
 Response via : Initial Calibration

Reviewed By :Krupa Patel 05/25/2023
 Supervised By :Mahesh Dadoda 05/25/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	180652	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	176342	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	91351	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	56226	36.418	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	72.840%	
7) Chloroethane-d5	1.535	69	55658	46.796	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	93.600%	
11) 1,1-Dichloroethene-d2	2.063	65	30246	41.777	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	83.560%	
21) 2-Butanone-d5	3.802	46	80834	98.260	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	98.260%	
24) Chloroform-d	4.259	84	119308	43.682	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	87.360%	
26) 1,2-Dichloroethane-d4	4.950	65	90589	49.083	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	98.160%	
32) Benzene-d6	4.969	84	208040	43.519	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	87.040%	
36) 1,2-Dichloropropane-d6	5.995	67	62338	43.582	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	87.160%	
41) Toluene-d8	7.249	98	187848	42.102	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	84.200%	
43) trans-1,3-Dichloroprop...	7.561	79	30921	43.088	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	86.180%	
47) 2-Hexanone-d5	8.034	63	47932	85.947	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	85.950%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	90961	42.399	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	84.800%	
66) 1,2-Dichlorobenzene-d4	11.567	152	78723	47.956	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	95.920%	
Target Compounds						
5) Vinyl chloride	1.288	62	2859m	2.507	ug/L	
20) cis-1,2-Dichloroethene	3.821	96	62642	54.769	ug/L	97
34) Trichloroethene	5.841	95	46504	41.082	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052423\
Data File : VV031222.D
Acq On : 24 May 2023 17:57
Operator : SY/MD
Sample : 02866-15DL 50X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBX90DL

Manual IntegrationsAPPROVED

Quant Time: May 25 01:16:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 25 01:11:46 2023
Response via : Initial Calibration

Reviewed By :Krupa Patel 05/25/2023
Supervised By :Mahesh Dadoda 05/25/2023

