

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052424\
 Data File : VW035754.D
 Acq On : 24 May 2024 18:03
 Operator : SY/MD
 Sample : P2577-11
 Misc : 5.37g/10.0mL/MSVOA_V/SOIL/VIAL A
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHBN9

Quant Time: May 24 23:49:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050624SMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 24 23:42:56 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.535	114	288508	25.000	ug/L	0.00
28) Chlorobenzene-d5		8.786	117	201127	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.188	152	39196	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.275	65	149748	33.581	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	134.320%	
7) Chloroethane-d5		1.529	69	118119	32.834	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	131.320%	
11) 1,1-Dichloroethene-d2		2.053	65	59179	29.977	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	119.920%#	
21) 2-Butanone-d5		3.854	46	43188	34.127	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.260%	
24) Chloroform-d		4.246	84	222979	25.997	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	104.000%	
26) 1,2-Dichloroethane-d4		4.944	65	120411	28.397	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	113.600%	
32) Benzene-d6		4.957	84	396318	35.997	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	144.000%#	
36) 1,2-Dichloropropane-d6		5.989	67	132338	38.950	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	155.800%#	
41) Toluene-d8		7.243	98	258924	26.425	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	105.680%	
43) trans-1,3-Dichloroprop...		7.561	79	31804	21.887	ug/L	0.01
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.560%	
47) 2-Hexanone-d5		8.040	63	19216	31.453	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	62.900%	
56) 1,1,2,2-Tetrachloroeth...		10.153	84	88758	27.110	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	108.440%	
66) 1,2-Dichlorobenzene-d4		11.561	152	32938	24.214	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.840%	
Target Compounds							
13) Acetone		2.147	43	10086	5.445	ug/L	Qvalue 73
16) Methylene chloride		2.442	84	26423	3.972	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052424\
Data File : VV035754.D
Acq On : 24 May 2024 18:03
Operator : SY/MD
Sample : P2577-11
Misc : 5.37g/10.0mL/MSVOA_V/SOIL/VIAL A
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHBN9

Quant Time: May 24 23:49:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050624SMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 24 23:42:56 2024
Response via : Initial Calibration

