

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052521\
 Data File : VV021391.D
 Acq On : 25 May 2021 16:25
 Operator : SY/MD
 Sample : M2496-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB434

Quant Time: May 26 07:53:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 26 07:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	251787	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	231582	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	126526	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	65193	6.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	122.40%
7) Chloroethane-d5	1.568	69	58954	5.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.00%
11) 1,1-Dichloroethene-d2	2.108	63	90646	4.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.80%
20) 2-Butanone-d5	3.905	46	211191	55.15	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.30%
24) Chloroform-d	4.352	84	153112	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.60%
26) 1,2-Dichloroethane-d4	5.037	65	72407	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.60%
32) Benzene-d6	5.053	84	308225	5.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.60%
36) 1,2-Dichloropropane-d6	6.072	67	96908	5.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.20%
41) Toluene-d8	7.320	98	281069	5.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.00%
43) trans-1,3-Dichloroprop...	7.628	79	35866	4.77	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.40%
46) 2-Hexanone-d5	8.091	63	185708	52.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.88%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	75918	5.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.20%
66) 1,2-Dichlorobenzene-d4	11.625	152	119124	5.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.20%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.117	96	12507	1.00	ug/L #	1
19) 1,1-Dichloroethane	3.198	63	4890	0.18	ug/L #	92
29) 1,1,1-Trichloroethane	4.609	97	71758	2.76	ug/L	97
42) Toluene	7.400	91	5344	0.08	ug/L	93
47) Tetrachloroethene	7.979	164	32350	2.14	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052521\
Data File : VV021391.D
Acq On : 25 May 2021 16:25
Operator : SY/MD
Sample : M2496-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB434

Quant Time: May 26 07:53:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 26 07:52:48 2021
Response via : Initial Calibration

