

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052521\
 Data File : VW021393.D
 Acq On : 25 May 2021 17:13
 Operator : SY/MD
 Sample : M2496-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB443

Manual Integrations
 APPROVED

MMDadoda
 5/26/2021 3:13:15 PM

Quant Time: May 26 07:53:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 26 07:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	253012	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	230742	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	124575	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	63373	5.92	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.40%
7) Chloroethane-d5	1.564	69	57837	5.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.20%
11) 1,1-Dichloroethene-d2	2.105	63	83637	4.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.40%
20) 2-Butanone-d5	3.902	46	199635	51.88	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.76%
24) Chloroform-d	4.352	84	148405	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
26) 1,2-Dichloroethane-d4	5.037	65	71704	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.00%
32) Benzene-d6	5.053	84	299088	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.80%
36) 1,2-Dichloropropane-d6	6.072	67	95372	5.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.00%
41) Toluene-d8	7.320	98	275631	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.20%
43) trans-1,3-Dichloroprop...	7.628	79	35559	4.75	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.00%
46) 2-Hexanone-d5	8.091	63	179701	50.93	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.86%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	74885	5.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.20%
66) 1,2-Dichlorobenzene-d4	11.625	152	115513	5.38	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.60%
Target Compounds						Qvalue
13) Acetone	2.201	43	9750m	4.49	ug/L	
25) Chloroform	4.381	83	18760	0.66	ug/L	93
38) Bromodichloromethane	6.516	83	48513	2.38	ug/L	96
42) Toluene	7.403	91	9260	0.14	ug/L	95
49) Dibromochloromethane	8.249	129	72283	4.82	ug/L	99
59) Bromoform	9.735	173	23621	2.62	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052521\
Data File : VV021393.D
Acq On : 25 May 2021 17:13
Operator : SY/MD
Sample : M2496-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB443

Manual Integrations
APPROVED

MMDadoda
5/26/2021 3:13:15 PM

Quant Time: May 26 07:53:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 26 07:52:48 2021
Response via : Initial Calibration

