

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052521\
 Data File : VW021397.D
 Acq On : 25 May 2021 18:49
 Operator : SY/MD
 Sample : M2496-20
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB460

Quant Time: May 26 07:54:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 26 07:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	245220	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	229433	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	124945	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	63550	6.13	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	122.60%	
7) Chloroethane-d5	1.565	69	57828	5.84	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	116.80%	
11) 1,1-Dichloroethene-d2	2.105	63	82226	4.18	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	83.60%	
20) 2-Butanone-d5	3.915	46	248228	66.56	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	133.12%#	
24) Chloroform-d	4.352	84	154121	5.36	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	107.20%	
26) 1,2-Dichloroethane-d4	5.037	65	76469	5.78	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	115.60%	
32) Benzene-d6	5.053	84	305186	5.53	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	110.60%	
36) 1,2-Dichloropropane-d6	6.072	67	96454	5.39	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	107.80%	
41) Toluene-d8	7.320	98	281823	5.57	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	111.40%	
43) trans-1,3-Dichloroprop...	7.629	79	37123	4.98	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	99.60%	
46) 2-Hexanone-d5	8.092	63	214849	61.24	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	122.48%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	83061	5.76	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	115.20%	
66) 1,2-Dichlorobenzene-d4	11.625	152	121727	5.65	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	113.00%	
Target Compounds						
13) Acetone	2.224	43	20466	9.73	ug/L	98
16) Methylene chloride	2.510	84	3694	0.18	ug/L	80
25) Chloroform	4.381	83	38603	1.41	ug/L	100
38) Bromodichloromethane	6.519	83	2528	0.12	ug/L	92
42) Toluene	7.400	91	14657	0.22	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052521\
Data File : VV021397.D
Acq On : 25 May 2021 18:49
Operator : SY/MD
Sample : M2496-20
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB460

Quant Time: May 26 07:54:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 26 07:52:48 2021
Response via : Initial Calibration

