

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052521\
 Data File : VV021404.D
 Acq On : 25 May 2021 21:37
 Operator : SY/MD
 Sample : M2496-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB440

Quant Time: May 26 07:55:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 26 07:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	234114	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	218105	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	117070	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	58662	5.92	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.40%
7) Chloroethane-d5	1.564	69	55234	5.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.80%
11) 1,1-Dichloroethene-d2	2.108	63	123058	6.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	131.00%#
20) 2-Butanone-d5	3.915	46	226585	63.64	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	127.28%
24) Chloroform-d	4.352	84	146918	5.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.00%
26) 1,2-Dichloroethane-d4	5.037	65	74186	5.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.40%
32) Benzene-d6	5.053	84	289245	5.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.20%
36) 1,2-Dichloropropane-d6	6.072	67	94776	5.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.40%
41) Toluene-d8	7.320	98	258779	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.60%
43) trans-1,3-Dichloroprop...	7.629	79	35188	4.97	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.40%
46) 2-Hexanone-d5	8.092	63	205794	61.70	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.40%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	79426	5.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.80%
66) 1,2-Dichlorobenzene-d4	11.625	152	114627	5.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.60%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.118	96	99616	8.58	ug/L	94
19) 1,1-Dichloroethane	3.191	63	46483	1.82	ug/L	98
22) cis-1,2-Dichloroethene	3.915	96	1784	0.11	ug/L	76
29) 1,1,1-Trichloroethane	4.613	97	73555	3.01	ug/L	100
34) Trichloroethene	5.928	95	2891	0.18	ug/L	91
42) Toluene	7.400	91	5416	0.09	ug/L	99
47) Tetrachloroethene	7.976	164	128413	9.01	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052521\
Data File : VV021404.D
Acq On : 25 May 2021 21:37
Operator : SY/MD
Sample : M2496-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB440

Quant Time: May 26 07:55:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 26 07:52:48 2021
Response via : Initial Calibration

