

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052521\
 Data File : VV021407.D
 Acq On : 25 May 2021 22:49
 Operator : SY/MD
 Sample : M2496-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB457

Quant Time: May 26 07:55:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 26 07:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	243973	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	224622	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	123538	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	58262	5.64	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	112.80%	
7) Chloroethane-d5	1.564	69	54137	5.50	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	110.00%	
11) 1,1-Dichloroethene-d2	2.108	63	127786	6.53	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	130.60%#	
20) 2-Butanone-d5	3.908	46	226029	60.92	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	121.84%	
24) Chloroform-d	4.352	84	143842	5.02	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.40%	
26) 1,2-Dichloroethane-d4	5.037	65	71777	5.45	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.00%	
32) Benzene-d6	5.053	84	285673	5.29	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.80%	
36) 1,2-Dichloropropane-d6	6.072	67	93499	5.34	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	106.80%	
41) Toluene-d8	7.320	98	261778	5.28	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.60%	
43) trans-1,3-Dichloroprop...	7.628	79	34352	4.71	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	94.20%	
46) 2-Hexanone-d5	8.091	63	202252	58.88	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	117.76%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	78328	5.55	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	111.00%	
66) 1,2-Dichlorobenzene-d4	11.625	152	114675	5.39	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.80%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.117	96	108006	8.92	ug/L	94
19) 1,1-Dichloroethane	3.191	63	20020	0.75	ug/L	98
22) cis-1,2-Dichloroethene	3.921	96	1764	0.11	ug/L	98
29) 1,1,1-Trichloroethane	4.609	97	138387	5.49	ug/L	98
34) Trichloroethene	5.928	95	2205	0.14	ug/L	96
42) Toluene	7.397	91	13522	0.21	ug/L	99
47) Tetrachloroethene	7.979	164	161198	10.98	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052521\
Data File : VV021407.D
Acq On : 25 May 2021 22:49
Operator : SY/MD
Sample : M2496-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB457

Quant Time: May 26 07:55:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 26 07:52:48 2021
Response via : Initial Calibration

