

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV052618\
 Data File : VV006049.D
 Acq On : 26 May 2018 12:16
 Operator : SY/MD
 Sample : VV0526WBL01
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 28 02:10:52 2018
 Quant Method : W:\HPCHEM1\MSVOA_V\METHOD\SOMVLM052518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon May 28 02:03:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	172455	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	157190	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	65225	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	60816	50.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.98%
7) Chloroethane-d5	1.56	69	40709	47.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.58%
11) 1,1-Dichloroethene-d2	2.13	63	85306	38.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.10%
21) 2-Butanone-d5	3.95	46	61794	89.57	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.57%
24) Chloroform-d	4.40	84	112513	49.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.12%
26) 1,2-Dichloroethane-d4	5.09	65	76238	51.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.88%
32) Benzene-d6	5.10	84	220437	52.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.50%
36) 1,2-Dichloropropane-d6	6.12	67	70257	52.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.88%
41) Toluene-d8	7.36	98	197495	51.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.18%
43) trans-1,3-Dichloropropene-	7.67	79	31140	50.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.90%
47) 2-Hexanone-d5	8.14	63	54062	103.65	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.65%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	94357	51.71	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.42%
64) 1,2-Dichlorobenzene-d4	11.68	152	72479	55.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.74%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV052618\
Data File : VV006049.D
Acq On : 26 May 2018 12:16
Operator : SY/MD
Sample : VV0526WBL01
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 28 02:10:52 2018
Quant Method : W:\HPCHEM1\MSVOA_V\METHOD\SOMVLM052518WMA.M
Quant Title : VOC Analysis
QLast Update : Mon May 28 02:03:29 2018
Response via : Initial Calibration

