

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV052618\
 Data File : VV006079.D
 Acq On : 27 May 2018 02:38
 Operator : SY/MD
 Sample : J3092-18 100X
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 36 Sample Multiplier: 1

Quant Time: May 28 04:35:44 2018
 Quant Method : W:\HPCHEM1\MSVOA_V\METHOD\SOMVLM052518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon May 28 02:03:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	162808	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	148809	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	57141	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	47635	41.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.78%
7) Chloroethane-d5	1.55	69	10833	13.47	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	26.94%#
11) 1,1-Dichloroethene-d2	2.12	63	70405	33.26	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.52%
21) 2-Butanone-d5	3.95	46	51722	79.42	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.42%
24) Chloroform-d	4.40	84	97096	45.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.60%
26) 1,2-Dichloroethane-d4	5.09	65	66825	47.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.52%
32) Benzene-d6	5.10	84	190806	48.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.46%
36) 1,2-Dichloropropane-d6	6.12	67	61483	48.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.88%
41) Toluene-d8	7.36	98	161342	44.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.18%
43) trans-1,3-Dichloropropene-	7.67	79	24290	41.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.96%
47) 2-Hexanone-d5	8.14	63	48428	98.08	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.08%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	85926	49.74	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.48%
64) 1,2-Dichlorobenzene-d4	11.68	152	60292	53.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV052618\
Data File : VV006079.D
Acq On : 27 May 2018 02:38
Operator : SY/MD
Sample : J3092-18 100X
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 36 Sample Multiplier: 1

Quant Time: May 28 04:35:44 2018
Quant Method : W:\HPCHEM1\MSVOA_V\METHOD\SOMVLM052518WMA.M
Quant Title : VOC Analysis
QLast Update : Mon May 28 02:03:29 2018
Response via : Initial Calibration

