

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052621\
 Data File : VV021421.D
 Acq On : 26 May 2021 13:26
 Operator : SY/MD
 Sample : M2496-21
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB461

Quant Time: May 27 02:04:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 27 02:03:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	238611	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	220408	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	125322	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	47286	4.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.60%
7) Chloroethane-d5	1.568	69	48623	5.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.00%
11) 1,1-Dichloroethene-d2	2.108	63	67556	3.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.60%
20) 2-Butanone-d5	3.902	46	225469	62.13	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.26%
24) Chloroform-d	4.352	84	137252	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	5.037	65	69898	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.60%
32) Benzene-d6	5.053	84	269990	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
36) 1,2-Dichloropropane-d6	6.076	67	89126	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	7.320	98	247243	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.60%
43) trans-1,3-Dichloroprop...	7.629	79	36710	5.13	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.60%
46) 2-Hexanone-d5	8.092	63	200001	59.34	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.68%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	75410	5.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.80%
66) 1,2-Dichlorobenzene-d4	11.625	152	113710	5.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.20%
Target Compounds						
3) Chloromethane	1.240	50	2789	0.20	ug/L	89
13) Acetone	2.198	43	4673	2.28	ug/L	84
25) Chloroform	4.378	83	11604	0.43	ug/L	93
38) Bromodichloromethane	6.516	83	20913	1.07	ug/L	96
42) Toluene	7.404	91	5429	0.09	ug/L	97
49) Dibromochloromethane	8.249	129	25162	1.76	ug/L	95
59) Bromoform	9.738	173	6903	0.76	ug/L #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052621\
Data File : VV021421.D
Acq On : 26 May 2021 13:26
Operator : SY/MD
Sample : M2496-21
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB461

Quant Time: May 27 02:04:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 27 02:03:26 2021
Response via : Initial Calibration

