

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052623\
 Data File : VW031247.D
 Acq On : 26 May 2023 12:26
 Operator : SY/MD
 Sample : VIBLK214
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK214

Quant Time: May 27 00:53:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat May 27 00:51:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	245304	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	281997	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	126622	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	110451	52.686	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 105.380%			
7) Chloroethane-d5	1.536	69	89416	55.365	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 110.740%			
11) 1,1-Dichloroethene-d2	2.063	65	41829	42.549	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 85.100%			
21) 2-Butanone-d5	3.796	46	171194	153.253	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 153.250%#			
24) Chloroform-d	4.253	84	226786	61.149	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 122.300%			
26) 1,2-Dichloroethane-d4	4.944	65	161570	64.470	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 128.940%#			
32) Benzene-d6	4.963	84	387916	50.743	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 101.480%			
36) 1,2-Dichloropropane-d6	5.992	67	133711	58.456	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 116.920%			
41) Toluene-d8	7.246	98	288176	40.389	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 80.780%			
43) trans-1,3-Dichloroprop...	7.558	79	60631	52.833	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 105.660%			
47) 2-Hexanone-d5	8.031	63	105868	118.708	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 118.710%			
56) 1,1,2,2-Tetrachloroeth...	10.159	84	203281	59.253	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 118.500%			
66) 1,2-Dichlorobenzene-d4	11.567	152	126973	55.803	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 111.600%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052623\
Data File : VV031247.D
Acq On : 26 May 2023 12:26
Operator : SY/MD
Sample : VIBLK214
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK214

Quant Time: May 27 00:53:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051923WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 27 00:51:17 2023
Response via : Initial Calibration

