

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052720\
 Data File : VV016233.D
 Acq On : 27 May 2020 17:16
 Operator : SY/MD
 Sample : L2754-15
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAXA1

Quant Time: May 28 04:58:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 28 01:17:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	78272	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	78701	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	33171	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	26414	5.25	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.00%
7) Chloroethane-d5	1.58	69	24713	5.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.00%
11) 1,1-Dichloroethene-d2	2.12	63	37369	3.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.20%
20) 2-Butanone-d5	3.94	46	75580	60.61	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.22%
24) Chloroform-d	4.38	84	53499	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.80%
26) 1,2-Dichloroethane-d4	5.06	65	29025	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.60%
32) Benzene-d6	5.08	84	103559	5.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.20%
36) 1,2-Dichloropropane-d6	6.10	67	33050	5.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.60%
41) Toluene-d8	7.34	98	92863	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
45) trans-1,3-Dichloropropene-	7.65	79	10499	4.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.40%
48) 2-Hexanone-d5	8.12	63	52355	50.78	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.56%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	22948	5.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	32789	5.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.40%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	1068	0.156 ug/L	# 46
18) trans-1,2-Dichloroethene	2.78	96	42904	8.492 ug/L	96
22) cis-1,2-Dichloroethene	3.94	96	166805	27.916 ug/L	99
34) Trichloroethene	5.94	95	42923	7.275 ug/L	96
49) Tetrachloroethene	8.00	164	49278	11.634 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052720\
Data File : VV016233.D
Acq On : 27 May 2020 17:16
Operator : SY/MD
Sample : L2754-15
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAXA1

Quant Time: May 28 04:58:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 28 01:17:22 2020
Response via : Initial Calibration

