

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052720\
 Data File : VV016238.D
 Acq On : 27 May 2020 19:17
 Operator : SY/MD
 Sample : L2755-12
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAXD6

Quant Time: May 28 06:00:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 28 01:17:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	82951	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	81996	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	36144	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	26252	4.93	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.60%
7) Chloroethane-d5	1.58	69	24252	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.60%
11) 1,1-Dichloroethene-d2	2.13	63	38006	3.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.00%
20) 2-Butanone-d5	3.94	46	73366	55.52	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.04%
24) Chloroform-d	4.38	84	54013	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
26) 1,2-Dichloroethane-d4	5.06	65	29995	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.80%
32) Benzene-d6	5.08	84	104102	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.60%
36) 1,2-Dichloropropane-d6	6.10	67	32714	5.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.20%
41) Toluene-d8	7.34	98	92375	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
45) trans-1,3-Dichloropropene-	7.65	79	10797	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
48) 2-Hexanone-d5	8.12	63	52079	48.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.96%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	23576	5.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	32498	5.38	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.60%

Target Compounds					Ovalue
34) Trichloroethene	5.94	95	8059	1.311 ug/L	97
49) Tetrachloroethene	8.00	164	12273	2.781 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052720\
Data File : VV016238.D
Acq On : 27 May 2020 19:17
Operator : SY/MD
Sample : L2755-12
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAXD6

Quant Time: May 28 06:00:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 28 01:17:22 2020
Response via : Initial Calibration

