

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052720\
 Data File : VV016244.D
 Acq On : 27 May 2020 22:09
 Operator : SY/MD
 Sample : L2766-01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BE4Y5

Quant Time: May 28 06:48:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 28 06:30:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	85859	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	84751	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	37667	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25263	4.58	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.60%
7) Chloroethane-d5	1.58	69	23976	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.00%
11) 1,1-Dichloroethene-d2	2.13	63	37119	3.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.60%
20) 2-Butanone-d5	3.94	46	76359	55.83	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.66%
24) Chloroform-d	4.38	84	52997	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
26) 1,2-Dichloroethane-d4	5.06	65	28949	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
32) Benzene-d6	5.08	84	101034	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	6.10	67	32203	4.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.20%
41) Toluene-d8	7.34	98	93932	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
45) trans-1,3-Dichloropropene-	7.65	79	10665	4.41	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.20%
48) 2-Hexanone-d5	8.12	63	51459	46.34	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.68%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	22481	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	32550	5.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.40%

Target Compounds

					Ovalue
13) Acetone	2.21	43	51505	60.831 ug/L	90
21) 2-Butanone	4.04	43	6791	4.509 ug/L	97
30) Cyclohexane	4.70	56	1771	0.164 ug/L	97
42) Toluene	7.41	91	6405	0.241 ug/L	95
64) 1,4-Dichlorobenzene	11.29	146	2982	0.255 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052720\
Data File : VV016244.D
Acq On : 27 May 2020 22:09
Operator : SY/MD
Sample : L2766-01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BE4Y5

Quant Time: May 28 06:48:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 28 06:30:54 2020
Response via : Initial Calibration

