

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052723\
 Data File : VV031252.D
 Acq On : 26 May 2023 17:36
 Operator : SY/MD
 Sample : VV0527WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK342

Quant Time: May 27 01:46:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052523WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat May 27 01:45:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	152144	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	148534	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	62345	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	71705	4.253	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.000%
7) Chloroethane-d5	1.536	69	58471	4.445	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.000%
11) 1,1-Dichloroethene-d2	2.060	65	29488	4.231	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.600%
20) 2-Butanone-d5	3.834	46	163721	45.112	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.220%
24) Chloroform-d	4.259	84	135874	4.063	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.200%
26) 1,2-Dichloroethane-d4	4.950	65	78454	4.749	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
32) Benzene-d6	4.970	84	258084	4.161	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.200%
36) 1,2-Dichloropropane-d6	5.998	67	91363	4.402	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.000%
41) Toluene-d8	7.249	98	203918	3.762	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.200%
43) trans-1,3-Dichloroprop...	7.561	79	25608	3.837	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.800%
46) 2-Hexanone-d5	8.037	63	119526	43.862	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.720%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	62809	4.288	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.800%
66) 1,2-Dichlorobenzene-d4	11.567	152	69499	4.751	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.000%
Target Compounds						Qvalue
16) Methylene chloride	2.452	84	19159	1.078	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052723\
Data File : VV031252.D
Acq On : 26 May 2023 17:36
Operator : SY/MD
Sample : VV0527WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK342

Quant Time: May 27 01:46:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052523WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat May 27 01:45:01 2023
Response via : Initial Calibration

