

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052820\
 Data File : VV016309.D
 Acq On : 29 May 2020 05:09
 Operator : SY/MD
 Sample : L2662-12
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: May 29 06:33:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 29 04:21:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	114753	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	108725	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	54698	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	28944	3.93	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.60%
7) Chloroethane-d5	1.58	69	21772	3.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	67.20%
11) 1,1-Dichloroethene-d2	2.12	63	42616	3.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.60%
20) 2-Butanone-d5	3.98	46	74053	40.51	ug/L	0.05
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.02%
24) Chloroform-d	4.38	84	59538	4.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.20%
26) 1,2-Dichloroethane-d4	5.06	65	32893	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.60%
32) Benzene-d6	5.07	84	123776	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.00%
36) 1,2-Dichloropropane-d6	6.10	67	37743	4.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.60%
41) Toluene-d8	7.34	98	114023	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.80%
45) trans-1,3-Dichloropropene-	7.65	79	11902	3.83	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.60%
48) 2-Hexanone-d5	8.12	63	72054	50.58	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.16%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	26451	4.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	42356	4.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.80%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.32	62	141764	14.106	ug/L	98
8) Chloroethane	1.55	64	21650	3.856	ug/L #	51
12) 1,1-Dichloroethene	2.13	96	1330	0.196	ug/L #	1
14) Carbon disulfide	2.31	76	906	0.043	ug/L #	75
18) trans-1,2-Dichloroethene	2.78	96	4180	0.564	ug/L	89
19) 1,1-Dichloroethane	3.21	63	2081	0.148	ug/L #	86
22) cis-1,2-Dichloroethene	3.94	96	84002	9.589	ug/L	94
27) 1,2-Dichloroethane	5.16	62	32021	3.289	ug/L	98
33) Benzene	5.12	78	144741	4.626	ug/L	100
34) Trichloroethene	5.94	95	3525	0.432	ug/L	97
35) Methylcyclohexane	6.10	83	9388	0.683	ug/L #	16
37) 1,2-Dichloropropane	6.09	63	4498	0.586	ug/L #	88
42) Toluene	7.41	91	36459	1.069	ug/L	99
43) 1,3,5-Trimethylbenzene	10.56	105	51565	1.660	ug/L	98
44) 1,2,4-Trimethylbenzene	10.94	105	30192	0.944	ug/L	97
46) trans-1,3-Dichloropropene	7.64	75	613	0.068	ug/L #	66
50) 2-Hexanone	8.12	43	10706	3.216	ug/L #	70

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052820\
 Data File : VV016309.D
 Acq On : 29 May 2020 05:09
 Operator : SY/MD
 Sample : L2662-12
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: May 29 06:33:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 29 04:21:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)


53) Chlorobenzene	8.90	112	165863	7.867	ug/L	99
54) Ethylbenzene	9.04	91	39130	1.029	ug/L	99
55) m,p-xylene	9.16	106	11182	0.784	ug/L	99
56) o-xylene	9.57	106	43465	3.198	ug/L	97
57) Styrene	9.57	104	2191	0.097	ug/L #	1
58) Isopropylbenzene	9.95	105	25631	0.688	ug/L	99
63) 1,3-Dichlorobenzene	11.21	146	4803	0.288	ug/L	90
64) 1,4-Dichlorobenzene	11.30	146	32259	1.902	ug/L	97
66) 1,2-Dichlorobenzene	11.67	146	162304	10.757	ug/L	100
67) 1,2-Dibromo-3-chloropropan	12.43	75	626	0.614	ug/L #	9
70) Naphthalene	13.53	128	4027832	177.734	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052820\
Data File : VV016309.D
Acq On : 29 May 2020 05:09
Operator : SY/MD
Sample : L2662-12
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: May 29 06:33:15 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 29 04:21:25 2020
Response via : Initial Calibration





#5

Vinyl chloride

Concen: 14.106 ug/L

RT: 1.32 min Scan# 71

Delta R.T. -0.00 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09


Instrument :
MSVOA_V
ClientSampled :

Tot Ion: 62 Resp: 141764

Ion Ratio Lower Upper

62 100

64 32.5 22.1 41.1



Abundance Ion 62.00 (61.70 to 62.70): VV0

Ion 64.10 (63.80 to 64.80): VV0

1.32



150000

100000

50000

0

Time--> 1.25 1.30 1.35 1.40



#8

Chloroethane

Concen: 3.856 ug/L

RT: 1.55 min Scan# 143

Delta R.T. -0.05 min

Lab File: VV016309.D


Acq: 29 May 2020 05:09

Tgt Ion: 64 Resp: 21650

Ion Ratio Lower Upper

64 100

66 5.1 22.9 42.5#



Abundance Ion 64.10 (63.80 to 64.80): VV0

Ion 66.05 (65.75 to 66.75): VV0

50000

40000


30000


20000

10000

0

Time--> 1.30 1.40 1.50 1.60









#12
 1,1-Dichloroethene
 Concen: 0.196 ug/L
 RT: 2.13 min Scan# 323
 Delta R.T. -0.01 min
 Lab File: VV016309.D
 Acq: 29 May 2020 05:09




Instrument :
 MSVOA_V
 ClientSampleId :


Tot Ion: 96 Resp: 1330
 Ion Ratio Lower Upper
 96 100
 61 361.0 124.9 231.9#
 63 1999.2 80.4 149.2#



#14
 Carbon disulfide
 Concen: 0.043 ug/L
 RT: 2.31 min Scan# 378
 Delta R.T. -0.01 min
 Lab File: VV016309.D
 Acq: 29 May 2020 05:09

Tgt Ion: 76 Resp: 906
 Ion Ratio Lower Upper
 76 100
 78 18.4 7.4 11.0#





#18

trans-1,2-Dichloroethene

Concen: 0.564 ug/L

RT: 2.78 min Scan# 524

Delta R.T. -0.01 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09

Instrument :
MSVOA_V
ClientSampled :


Tot Ion: 96 Resp: 4180

Ion Ratio Lower Upper




96 100

61 157.2 99.8 185.4

98 69.7 44.0 81.6



Abundance Ion 96.00 (95.70 to 96.70): VV0
Ion 61.05 (60.75 to 61.75): VV0
Ion 97.95 (97.65 to 98.65): VV0



#19

1,1-Dichloroethane

Concen: 0.148 ug/L

RT: 3.21 min Scan# 660

Delta R.T. 0.00 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09


Tgt Ion: 63 Resp: 2081

Ion Ratio Lower Upper



63 100


65 40.5 22.3 41.5

83 18.1 9.4 17.4#



Abundance Ion 63.10 (62.80 to 63.80): VV0
Ion 65.10 (64.80 to 65.80): VV0
Ion 83.05 (82.75 to 83.75): VV0






#22




cis-1,2-Dichloroethene
 Concen: 9.589 ug/L
 RT: 3.94 min Scan# 885
 Delta R.T. -0.01 min
 Lab File: VV016309.D
 Acq: 29 May 2020 05:09

Instrument :
 MSVOA_V
 ClientSampled :

Tot Ion: 96 Resp: 84002
 Ion Ratio Lower Upper
 96 100
 61 135.2 99.7 185.1
 68 0.0 0.0 0.0




Abundance Ion 96.00 (95.70 to 96.70): VV0
 Ion 61.05 (60.75 to 61.75): VV0
 Ion 68.15 (67.85 to 68.85): VV0





#27


1,2-Dichloroethane
 Concen: 3.289 ug/L
 RT: 5.16 min Scan# 1266
 Delta R.T. 0.00 min
 Lab File: VV016309.D
 Acq: 29 May 2020 05:09

Tgt Ion: 62 Resp: 32021
 Ion Ratio Lower Upper
 62 100
 98 7.8 5.7 8.5



Abundance Ion 62.00 (61.70 to 62.70): VV0
 Ion 98.05 (97.75 to 98.75): VV0





#33

Benzene

Concen: 4.626 ug/L

RT: 5.12 min Scan# 1254


Delta R.T. -0.01 min

Lab File: VV016309.D




Acq: 29 May 2020 05:09

Instrument :
MSVOA_V
ClientSampled :

Tgt Ion: 78 Resp: 144741



Abundance Ion 78.10 (77.80 to 78.80): VV0



#34

Trichloroethene

Concen: 0.432 ug/L

RT: 5.94 min Scan# 1508

Delta R.T. 0.00 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09

Tgt Ion: 95 Resp: 3525




Ion	Ratio	Lower	Upper
95	100		
97	55.7	43.6	81.0
132	88.9	63.9	118.7
130	93.4	65.2	121.0


Abundance Ion 95.00 (94.70 to 95.70): VV0

Ion 96.95 (96.65 to 97.65): VV0

Ion 131.95 (131.65 to 132.65): VV0

Ion 129.95 (129.65 to 130.65): VV0





#35

Methylvlcyclohexane

Concen: 0.683 ug/L

RT: 6.10 min Scan# 1557

Delta R.T. -0.06 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09

Instrument :

MSVOA_V

ClientSampleId :


Tot Ion: 83 Resp: 9388

Ion Ratio Lower Upper

83 100

55 0.0 65.4 98.0#




98 1.1 39.3 58.9#



Abundance Ion 83.15 (82.85 to 83.85): VV016309.D (-1482) (-)

Ion 55.10 (54.80 to 55.80): VV016309.D (-1482) (-)

Ion 98.15 (97.85 to 98.85): VV016309.D (-1482) (-)



#37

1,2-Dichloropropane

Concen: 0.586 ug/L

RT: 6.09 min Scan# 1556

Delta R.T. -0.11 min

Lab File: VV016309.D


Acq: 29 May 2020 05:09

Tgt Ion: 63 Resp: 4498

Ion Ratio Lower Upper



63 100


112 0.0 3.1 4.7#



Abundance Ion 63.10 (62.80 to 63.80): VV016309.D (-1496) (-)

Ion 112.10 (111.80 to 112.80): VV016309.D (-1496) (-)





#42

Toluene

Concen: 1.069 ug/L

RT: 7.41 min Scan# 1966

Delta R.T. 0.00 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09


Instrument :
MSVOA_V
ClientSampleId :

Tot Ion: 91 Resp: 36459

Ion Ratio Lower Upper

91 100

92 57.6 40.9 76.0



Abundance Ion 91.15 (90.85 to 91.85): VV016290.D (-1952) (-)

Ion 92.15 (91.85 to 92.85): VV016290.D (-1952) (-)

7.41

20000

15000



10000

5000

0

Time-->

7.35 7.40 7.45 7.50



#43

1,3,5-Trimethylbenzene

Concen: 1.660 ug/L

RT: 10.56 min Scan# 2946

Delta R.T. 0.00 min

Lab File: VV016309.D


Acq: 29 May 2020 05:09

Tgt Ion: 105 Resp: 51565

Ion Ratio Lower Upper

105 100

120 48.0 37.4 56.0



Abundance Ion 105.00 (104.70 to 105.70): VV016290.D (-2933) (-)

Ion 120.00 (119.70 to 120.70): VV016290.D (-2933) (-)

10.56

40000

30000


20000


10000

0

Time-->

10.50 10.55 10.60









#44
 1,2,4-Trimethylbenzene
 Concen: 0.944 ug/L
 RT: 10.94 min Scan# 3062
 Delta R.T. 0.00 min
 Lab File: VV016309.D
 Acq: 29 May 2020 05:09




Instrument :
 MSVOA_V
 ClientSampled :


Tot Ion:105 Resp: 30192
 Ion Ratio Lower Upper
 105 100
 120 45.4 34.6 52.0



#46
 trans-1,3-Dichloropropene
 Concen: 0.068 ug/L
 RT: 7.64 min Scan# 2038
 Delta R.T. -0.03 min
 Lab File: VV016309.D
 Acq: 29 May 2020 05:09

Tgt Ion: 75 Resp: 613
 Ion Ratio Lower Upper
 75 100
 77 49.4 21.7 40.3#









#50
2-Hexanone
Concen: 3.216 ug/L
RT: 8.12 min Scan# 2186
Delta R.T. -0.05 min
Lab File: VV016309.D
Acq: 29 May 2020 05:09




Instrument :
MSVOA_V
ClientSampleId :


Tot Ion	Ion	Ratio	Lower	Upper
43	100			
58	23.3	41.4	62.0	
57	18.2	15.0	22.4	
100	0.0	9.0	13.4	



#53
Chlorobenzene
Concen: 7.867 ug/L
RT: 8.90 min Scan# 2430
Delta R.T. 0.00 min
Lab File: VV016309.D
Acq: 29 May 2020 05:09

Tgt Ion	Ion	Ratio	Lower	Upper
112	100			
114	32.7	22.3	41.3	
77	62.7	50.7	76.1	





#54

Ethylbenzene

Concen: 1.029 ug/L

RT: 9.04 min Scan# 2471

Delta R.T. 0.00 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09


Instrument :
MSVOA_V
ClientSampled :

Tot Ion: 91 Resp: 39130




Ion Ratio Lower Upper

91 100

106 29.7 21.0 39.0



Abundance Ion 91.15 (90.85 to 91.85): VV016309.D (-2478) (-)



#55

m,p-xylene

Concen: 0.784 ug/L

RT: 9.16 min Scan# 2510

Delta R.T. 0.00 min

Lab File: VV016309.D


Acq: 29 May 2020 05:09

Tgt Ion: 106 Resp: 11182



Ion Ratio Lower Upper


106 100

91 213.0 148.1 275.0



Abundance Ion 106.15 (105.85 to 106.85): VV016309.D (-2517) (-)





#56

o-xylene

Concen: 3.198 ug/L

RT: 9.57 min Scan# 2636

Delta R.T. 0.00 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09


Instrument :
MSVOA_V
ClientSampled :

Tot Ion:106 Resp: 43465

Ion Ratio Lower Upper




106 100

91 221.2 158.8 294.8



Abundance Ion 106.15 (105.85 to 106.85): V

Ion 91.15 (90.85 to 91.85): VV0



#57

Styrene

Concen: 0.097 ug/L

RT: 9.57 min Scan# 2636

Delta R.T. -0.02 min

Lab File: VV016309.D


Acq: 29 May 2020 05:09

Tgt Ion:104 Resp: 2191

Ion Ratio Lower Upper



104 100


78 366.0 33.2 61.7#



Abundance Ion 104.10 (103.80 to 104.80): V

Ion 78.10 (77.80 to 78.80): VV0





#58

Isopropylbenzene

Concen: 0.688 ug/L

RT: 9.95 min Scan# 2757

Delta R.T. 0.00 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09

Instrument :

MSVOA_V

ClientSampleId :


Tot Ion:105 Resp: 25631

Ion Ratio Lower Upper

105 100

120 25.6 20.9 31.3




77 17.3 13.9 20.9



Abundance Ion 105.05 (104.75 to 105.75): V

Ion 120.10 (119.80 to 120.80): V

Ion 77.10 (76.80 to 77.80): V



#63

1,3-Dichlorobenzene

Concen: 0.288 ug/L

RT: 11.21 min Scan# 3146

Delta R.T. 0.00 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09


Tgt Ion:146 Resp: 4803

Ion Ratio Lower Upper

146 100

111 51.4 29.7 55.1



148 71.1 45.9 85.3




Abundance Ion 145.95 (145.65 to 146.65): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V





#64

1,4-Dichlorobenzene

Concen: 1.902 ug/L

RT: 11.30 min Scan# 3174

Delta R.T. 0.00 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09

Instrument :

MSVOA_V

ClientSampleId :


Tot Ion:146 Resp: 32259

Ion Ratio Lower Upper

146 100

111 40.9 28.3 52.7




148 64.7 43.3 80.3



Abundance Ion 145.95 (145.65 to 146.65): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#66

1,2-Dichlorobenzene

Concen: 10.757 ug/L

RT: 11.67 min Scan# 3289

Delta R.T. 0.00 min

Lab File: VV016309.D

Acq: 29 May 2020 05:09


Tgt Ion:146 Resp: 162304

Ion Ratio Lower Upper

146 100

111 44.5 35.7 53.5



148 64.0 51.3 76.9




Abundance Ion 145.95 (145.65 to 146.65): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V









#67
 1,2-Dibromo-3-chloropropane
 Concen: 0.614 ug/L
 RT: 12.43 min Scan# 3527
 Delta R.T. -0.02 min
 Lab File: VV016309.D
 Acq: 29 May 2020 05:09

Instrument :
 MSVOA_V
 ClientSampleId :

Tot Ion: 75 Resp: 626
 Ion Ratio Lower Upper
 75 100
 155 0.0 56.2 84.4#
 157 0.0 71.9 107.9#



#70
 Naphthalene
 Concen: 177.734 ug/L
 RT: 13.53 min Scan# 3868
 Delta R.T. 0.00 min
 Lab File: VV016309.D
 Acq: 29 May 2020 05:09

Tgt Ion: 128 Resp: 4027832
 Ion Ratio Lower Upper
 128 100
 129 11.2 8.6 13.0
 127 13.3 10.6 15.8

