

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052820\
 Data File : VV016273.D
 Acq On : 28 May 2020 13:44
 Operator : SY/MD
 Sample : L2755-13DL 5X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAXD7DL

Quant Time: May 29 02:06:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 29 01:10:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.65	114	85325	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	83999	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	37805	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	27896	5.09	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.80%
7) Chloroethane-d5	1.58	69	21581	4.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.60%
11) 1,1-Dichloroethene-d2	2.13	63	38858	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.60%
20) 2-Butanone-d5	3.94	46	68490	50.39	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.78%
24) Chloroform-d	4.38	84	50810	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.06	65	28081	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.20%
32) Benzene-d6	5.08	84	100694	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.80%
36) 1,2-Dichloropropane-d6	6.10	67	30691	4.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.40%
41) Toluene-d8	7.34	98	89981	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
45) trans-1,3-Dichloropropene-	7.65	79	10599	4.42	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.40%
48) 2-Hexanone-d5	8.12	63	51993	47.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.48%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	21035	4.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.40%
65) 1,2-Dichlorobenzene-d4	11.65	152	31670	5.02	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.40%

Target Compounds

					Ovalue
13) Acetone	2.22	43	2118	2.517 ug/L	75
15) Methyl Acetate	2.45	43	1178	0.548 ug/L	# 85
16) Methylene chloride	2.53	84	4902	0.850 ug/L	94
25) Chloroform	4.41	83	2644	0.239 ug/L	92
34) Trichloroethene	5.94	95	6185	0.982 ug/L	94
49) Tetrachloroethene	8.00	164	94652	20.936 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052820\
Data File : VV016273.D
Acq On : 28 May 2020 13:44
Operator : SY/MD
Sample : L2755-13DL 5X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAXD7DL

Quant Time: May 29 02:06:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 29 01:10:53 2020
Response via : Initial Calibration

