

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052820\
 Data File : VV016280.D
 Acq On : 28 May 2020 16:34
 Operator : SY/MD
 Sample : L2755-17
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAXD9

Quant Time: May 29 02:35:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 29 01:10:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	88312	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	85362	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	38214	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22856	4.03	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.60%
7) Chloroethane-d5	1.58	69	21974	4.41	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.20%
11) 1,1-Dichloroethene-d2	2.13	63	34231	3.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.20%
20) 2-Butanone-d5	3.94	46	77126	54.82	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.64%
24) Chloroform-d	4.38	84	51389	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.00%
26) 1,2-Dichloroethane-d4	5.06	65	30325	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.40%
32) Benzene-d6	5.08	84	101482	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	6.10	67	33187	5.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.40%
41) Toluene-d8	7.34	98	91127	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
45) trans-1,3-Dichloropropene-	7.65	79	10856	4.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.00%
48) 2-Hexanone-d5	8.12	63	56623	50.63	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.26%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	23791	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.40%
65) 1,2-Dichlorobenzene-d4	11.65	152	32623	5.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.20%

Target Compounds

					Ovalue
18) trans-1,2-Dichloroethene	2.78	96	943	0.165 ug/L	96
22) cis-1,2-Dichloroethene	3.94	96	1901	0.282 ug/L	90
25) Chloroform	4.41	83	4386	0.383 ug/L	99
33) Benzene	5.13	78	4660	0.190 ug/L	100
34) Trichloroethene	5.94	95	5546	0.867 ug/L	87
42) Toluene	7.42	91	2341	0.087 ug/L	92
49) Tetrachloroethene	8.00	164	83686	18.215 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052820\
Data File : VV016280.D
Acq On : 28 May 2020 16:34
Operator : SY/MD
Sample : L2755-17
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAXD9

Quant Time: May 29 02:35:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 29 01:10:53 2020
Response via : Initial Calibration

