

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052820\
 Data File : VV016292.D
 Acq On : 28 May 2020 22:16
 Operator : SY/MD
 Sample : L2754-15DL 4X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAXA1DL

Quant Time: May 29 04:35:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 29 04:21:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	98630	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	96773	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	42192	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	27692	4.37	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.40%
7) Chloroethane-d5	1.58	69	21965	3.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.00%
11) 1,1-Dichloroethene-d2	2.13	63	38746	3.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.20%
20) 2-Butanone-d5	3.95	46	71363	45.42	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.84%
24) Chloroform-d	4.38	84	49548	3.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.60%
26) 1,2-Dichloroethane-d4	5.06	65	28307	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.80%
32) Benzene-d6	5.08	84	101261	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.40%
36) 1,2-Dichloropropane-d6	6.10	67	30997	4.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.60%
41) Toluene-d8	7.34	98	90420	3.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.20%
45) trans-1,3-Dichloropropene-	7.65	79	10369	3.75	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.00%
48) 2-Hexanone-d5	8.12	63	52714	41.58	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.16%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	21709	3.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	31940	4.53	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.60%

Target Compounds

					Ovalue
18) trans-1,2-Dichloroethene	2.78	96	9796	1.539 ug/L	90
22) cis-1,2-Dichloroethene	3.94	96	39895	5.299 ug/L	97
34) Trichloroethene	5.94	95	10101	1.392 ug/L	98
49) Tetrachloroethene	8.00	164	11068	2.125 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052820\
Data File : VV016292.D
Acq On : 28 May 2020 22:16
Operator : SY/MD
Sample : L2754-15DL 4X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAXA1DL

Quant Time: May 29 04:35:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 29 04:21:25 2020
Response via : Initial Calibration

