

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052820\
 Data File : VV016282.D
 Acq On : 28 May 2020 17:22
 Operator : SY/MD
 Sample : L2755-20
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: May 29 02:44:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 29 01:10:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	78862	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	78941	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	34921	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	26923	5.31	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.20%
7) Chloroethane-d5	1.58	69	20455	4.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.00%
11) 1,1-Dichloroethene-d2	2.12	63	37833	3.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.60%
20) 2-Butanone-d5	3.94	46	66892	53.24	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.48%
24) Chloroform-d	4.38	84	47781	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
26) 1,2-Dichloroethane-d4	5.06	65	27181	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.80%
32) Benzene-d6	5.08	84	95972	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
36) 1,2-Dichloropropane-d6	6.10	67	29271	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.34	98	84316	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
45) trans-1,3-Dichloropropene-	7.65	79	9654	4.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.60%
48) 2-Hexanone-d5	8.12	63	48711	47.10	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.20%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	20914	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	29713	5.09	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052820\
Data File : VV016282.D
Acq On : 28 May 2020 17:22
Operator : SY/MD
Sample : L2755-20
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: May 29 02:44:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 29 01:10:53 2020
Response via : Initial Calibration

