


Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV052820\
 Data File : VV016264.D
 Acq On : 28 May 2020 09:40
 Operator : SY/MD
 Sample : BFB33
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB33

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri May 29 04:21:25 2020



AutoFind: Scans 2800, 2801, 2802; Background Corrected with Scan 2788

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	12605	PASS
75	95	30	80	54.5	31421	PASS
95	95	100	100	100.0	57602	PASS
96	95	5	9	7.1	4089	PASS
173	174	0.00	2	1.4	583	PASS
174	95	50	120	69.9	40242	PASS
175	174	5	9	7.6	3044	PASS
176	174	95	101	95.5	38429	PASS
177	176	5	9	6.4	2460	PASS

Scan 2788 (10.055 min): VV016264.D
BFB33

m/z	Abundance
39.90	131.0
44.00	250.0

Instrument :
MSVOA_V
ClientSampleId :
BFB33

m/z	Abundance
36.00	585.0
37.00	3167.0
38.00	2912.0
39.00	1391.0
39.90	197.0
44.00	691.0
45.00	629.0
46.10	108.0
46.90	673.0
48.00	379.0
49.00	2911.0
50.00	12385.0
51.00	3649.0
51.90	254.0
55.00	177.0
55.90	1008.0
57.00	1790.0
58.00	100.0
59.90	625.0
61.00	3100.0
62.00	3257.0
63.00	2443.0
64.00	347.0
64.90	286.0
67.00	218.0
68.00	6586.0
69.00	6602.0
70.00	511.0
72.00	295.0
73.00	2892.0
74.00	10611.0
75.00	30480.0
76.00	2505.0
77.00	441.0
77.90	290.0
78.90	2231.0
79.80	528.0
80.90	2327.0
81.90	531.0
85.90	118.0
86.90	2275.0
87.90	2038.0
90.90	255.0
92.00	1709.0
92.90	2784.0
94.00	6948.0
95.00	54984.0
96.00	4128.0
96.90	151.0
103.90	381.0
104.80	116.0
105.80	285.0
115.70	239.0
116.90	493.0
117.80	267.0
118.80	312.0
127.90	181.0
129.80	212.0
134.80	153.0
140.90	684.0
141.90	113.0
142.80	758.0
145.70	125.0
147.90	129.0
154.80	137.0
172.00	454.0
173.00	563.0
173.90	37648.0
175.00	2964.0
175.90	35760.0
176.90	2333.0

Instrument :
MSVOA_V
ClientSampleId :
BFB33

m/z	Abundance
36.00	652.0
37.00	3486.0
38.00	3214.0
39.00	1384.0
39.90	212.0
44.00	666.0
45.00	691.0
47.00	740.0
47.90	416.0
49.00	3041.0
50.00	12885.0
51.00	3867.0
51.90	259.0
54.90	177.0
56.00	1032.0
57.00	1899.0
57.90	112.0
60.00	580.0
61.00	3308.0
62.00	3529.0
63.00	2521.0
63.90	327.0
64.90	322.0
66.90	215.0
68.00	7095.0
69.00	7098.0
70.00	537.0
72.00	341.0
73.00	3223.0
74.00	11517.0
75.00	32440.0
76.00	2596.0
77.00	447.0
77.90	303.0
78.90	2382.0
79.80	618.0
80.90	2628.0
81.90	559.0
87.00	2382.0
88.00	2264.0
90.90	247.0
92.00	1903.0
92.90	2904.0
94.00	7603.0
95.00	59520.0
96.00	4240.0
96.90	130.0
103.80	398.0
104.80	116.0
105.90	300.0
115.70	291.0
116.90	517.0
117.80	291.0
118.80	341.0
128.00	209.0
128.90	107.0
129.80	239.0
134.80	146.0
140.90	696.0
142.00	120.0
142.80	747.0
145.70	122.0
147.90	154.0
154.80	155.0
171.90	458.0
173.00	602.0
173.90	41664.0
175.00	3108.0
175.90	39648.0
176.90	2533.0

Instrument :
MSVOA_V
ClientSampleId :
BFB33

m/z	Abundance
36.00	650.0
37.00	3411.0
38.00	3183.0
39.00	1222.0
39.90	235.0
44.00	566.0
45.00	650.0
47.00	702.0
47.90	443.0
49.00	2848.0
50.00	12546.0
51.00	3800.0
51.90	227.0
54.90	207.0
56.00	911.0
57.00	1891.0
57.90	111.0
60.00	550.0
61.00	3150.0
62.00	3396.0
63.00	2390.0
63.90	298.0
64.90	334.0
66.90	198.0
68.00	6837.0
69.00	6793.0
70.00	469.0
71.90	371.0
73.00	3137.0
74.00	11372.0
75.00	31344.0
76.00	2589.0
77.00	452.0
77.90	312.0
78.90	2253.0
79.80	615.0
80.90	2564.0
81.90	533.0
87.00	2226.0
88.00	2230.0
90.90	246.0
92.00	1888.0
92.90	2780.0
94.00	7400.0
95.00	58304.0
96.00	3899.0
96.90	105.0
103.80	375.0
104.70	122.0
105.90	309.0
115.70	288.0
116.90	459.0
117.80	260.0
118.80	391.0
128.00	228.0
128.90	135.0
129.80	243.0
134.70	140.0
140.90	652.0
142.00	102.0
142.80	731.0
145.80	103.0
147.90	148.0
154.90	144.0
156.90	104.0
171.80	434.0
172.90	584.0
173.90	41416.0
174.90	3060.0
175.90	39880.0
176.90	2516.0
177.80	115.0

Instrument :
MSVOA_V
ClientSampleId :
BFB33