

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052920\
 Data File : VV016318.D
 Acq On : 29 May 2020 16:21
 Operator : SY/MD
 Sample : L2662-02DL 20X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9B06DL

Quant Time: May 30 01:15:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 30 00:00:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	99865	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	92721	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	42243	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	29887	4.66	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.20%
7) Chloroethane-d5	1.58	69	24213	4.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.00%
11) 1,1-Dichloroethene-d2	2.12	63	42168	3.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.00%
20) 2-Butanone-d5	3.96	46	83742	52.64	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.28%
24) Chloroform-d	4.38	84	58235	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
26) 1,2-Dichloroethane-d4	5.06	65	32395	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.80%
32) Benzene-d6	5.07	84	116473	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.40%
36) 1,2-Dichloropropane-d6	6.10	67	35664	5.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.40%
41) Toluene-d8	7.34	98	106128	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
45) trans-1,3-Dichloropropene-	7.65	79	11725	4.43	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.60%
48) 2-Hexanone-d5	8.12	63	65293	53.75	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.50%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	24990	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	36562	5.18	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.60%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	4477	0.512 ug/L	90
18) trans-1,2-Dichloroethene	2.77	96	1194	0.185 ug/L	87
22) cis-1,2-Dichloroethene	3.94	96	28418	3.728 ug/L	99
34) Trichloroethene	5.94	95	69779	10.039 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052920\
Data File : VV016318.D
Acq On : 29 May 2020 16:21
Operator : SY/MD
Sample : L2662-02DL 20X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9B06DL

Quant Time: May 30 01:15:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 30 00:00:26 2020
Response via : Initial Calibration

